



EASYCAL 4.0

Test record

BRAND

YOEL NAIM LTD
verification lab
POB 261
74101 NESS ZIONA

Instrument data:		Results from gravimetric tests: EX			
Instrument:	HandyStep electronic	Test volume: (µl)	500	250	50
No.	500 µl Digital 10 G 27539	Value: (mg)	497.86	248.93	49.79
Thermometer:	LSW Physics	X 1	496.2	248.4	49.6
No.	1	X 2	496	248.8	49
Balance:	A & D HR	X 3	497.1	248.5	49.8
No.	36	X 4	497.1	248.5	49.6
Temperature:	26.18 °C / 79.12 °F	X 5	497.3	248.4	49.7
Relative humidity	50% ± 15%	X 6	497.4	248.6	49.6
Correction factor z:	1.00430	X 7	497.5	248.6	49.6
Atmosph. pressure abs(hPa):	1000	X 8	497.3	248	49.7
Note		X 9	497.2	248.5	49.8
		X 10	496.8	248.6	49.6
		Data analysis:			
		Test volume: µl	500	250	50
		V Mean µl	499.13	249.56	49.81
		A(%) Actual	-0.17	-0.18	-0.37
		CV(%) Actual	0.10	0.08	0.46
		A(%) Nominal	.9	.9	.9
		CV(%) Nominal	.25	.5	1
		Result A	ok	ok	ok
		Result CV	ok	ok	ok

The test was carried out according to:

ISO 8655-2002

Next test

11.2012

Result:

Gravimetric test ok

Test date:

10.05.2012

Test Operator:

zvika

Signature: